

TENDER NOTICE

2nd Call

Malda, Dated.....

27.06.2015

Sealed Tenders are invited from the reputed Dealers/Suppliers for supplying 300 MA X-Ray Machine as per specification enclosed for the purpose of proper functioning related to medical treatment at Matri Sadan Englishbazar Municipality in the interest of public service.

The tender paper accompanied with an Earnest Money of an amount of Rs. 25,000/- (Rupees Twenty-Five Thousand) only in the form of Bank Draft on any scheduled Bank pledged in favour of Chairman, Englishbazar Municipality should be dropped at the Tender Box lying at the General Section of Englishbazar Municipality, in person or by Speed Post on & from7.6.15.....

Last Time & Date of receipt of Tender papers is 3 p.m. on19.5.15.....

Conditional, Illegible, ambiguous Tender Paper/s will be out rightly rejected.

Rate against each of the Equipments should be quoted in specific in words and figures and should be inclusive of all Taxes, Vat etc. No other charges shall be allowed extra.

Tender papers so deposited will be opened on next Office day at 12 noon or any other subsequent day convenient to the undersigned in presence of the willing Tenderers.


Payment shall be made after receipt/installation of each equipment subject to fulfillment of terms in regard to quality and submission of bill.

Successful Tenderer must supply the quoted equipment within 10 days from the date of receipt of the order from the undersigned.

Successful Tenderers will have to enter in to Agreement with Chairman, Englishbazar Municipality depicting Terms & Conditions in regard to supply & payment etc.

The undersigned reserves the right to reject, approve, prefer any Tender without assigning any reason whatsoever in any stage.

The undersigned also reserves the right to re-invite the Tender at his sole discretion.


Chairman

Englishbazar Municipality, Malda

Memo No. 638 | VIII. 11 | 15-16 (7) Dated 27.06.2015

Copy forwarded for information & taking necessary action to:-

1. The District Magistrate, Malda
2. The Executive Officer, Englishbazar Municipality, Malda
3. The Finance Officer, Englishbazar Municipality, Malda
4. The U.P. Englishbazar Municipality, Malda with director to upload all the matters in the Website of the Englishbazar Municipality, Malda
5. The Accountant, Englishbazar Municipality, Malda
6. The Dealing Assistant, Matrisadan, Englishbazar Municipality, Malda
7. In-Charge General Section, Englishbazar Municipality, Malda with direct to display at N.B. of the ULB.


Chairman

Englishbazar Municipality, Malda

Memorandum - 638 (VIII) - 11 | 15-16 J-27.06.2015

TECHNICAL SPECIFICATIONS 300MA X-RAY MACHINE

X-RAY MACHINE	300mA. 125 KVP Full Wave Solid State Silicon Rectified X-ray Generator for Radiography suitable for single Tube Operation as per IS:7620 with latest amendments.
TIMER	The exposure timer should be digital & from 0.02 sec to 5 sec.
CONTROL	Consisting of On & Off Switches, with Voltmeter, MA meter, Quick trip-overload circuit breaker (Automatic safety system to block unwanted exposure factors beyond Tube Rating), Voltage Compensator, Major & Minor KV selector, Tube overload indicator provided, Space charge compensator, technique selector. Digital Display of KVP, & mAs, Anode braking Device. Electronic overload for protection of tube H.V. Cables & H.T. Tank. Ready & X-Ray Switch on Control.
Static balancer	30 KVA capable of converting 3-phase in range 365-400 VAC with 0.2 ohms resistance at 50-60Hz to single phase 230 VAC/50Hz
HV TRANSFORMER	Compact Heavy Duty Transformer comprising HV Silicon Rectifiers, HT Transformer, Filament Transformer, bushings all immersed in oil.
ACCESSORIES	Hand Switch with Flexible long cord enables the operator to keep away from the Radiation area during exposure. - Aluminum Filter. - Light Beam Diaphragm.
POWER SUPPLY REQUIREMENT	440V ± 10% V AC, 50 Hz, Three Phase -
TUBE UNIT	One No. 20/40 Rotating Anode Tube (BEL / TOSHIBA or equivalent) with Dual Focus (Large & Small).
HV CABLE	One Pair of 10 meter High Voltage Cables : Sleeving - Straight.
STAND	Floor to Ceiling Stand & with Counter Balanced Tube Head (Rotatable+180 Degree), 360 Degree Rotatable; mounted on Floor Ceiling Rails for convenient movements. It should have all necessary locks. The column should be light in weight. Protection for the rope failure should be provided with the column stand.
TABLE	5 Position Table Hand tilt; 15 Degree trendelenburg to vertical, Motorised Bucky consisting of 8:1, 103 lines/inch Imported Grid Size 17 1/4" X 18 7/8"; Stainless Steel Cassette Tray; Compression Band and Foot Rest.
VERTICAL BUCKY	Vertical Bucky Stand with moving Grid of Ratio 8:1, 103 lines/inch. The Bucky moves up & down & are equipped with stainless steel cassette. This stand is floor-mounted type & can accommodate cassettes up to 14" X 17".


Chairman